

# Pulsar PRO



## NITRILE EXAMINATION GLOVES



NON STERILE



SINGLE USE ONLY



AMBIDEXTROUS



### PREMIUM QUALITY. PRACTICAL CHOICE.

Medical Grade Chemo-Rated.

4+ mils Thickness.

Fentanyl Tested.

Easy on and off. Worry-free protection.



ASTM  
6319

ASTM  
6978



MANUFACTURED IN MALAYSIA

#### Features

- 100% Nitrile (Acrylonitrile-butadiene)
- Powder Free
- Smooth and Beaded cuff
- Finger Textured
- Ambidextrous
- Straight Fingers
- Comfortable Fit
- Cleared for use in Chemotherapy
- Fentanyl Tested
- Color: Blue

#### Packaging

- 100 Gloves per Box by Weight
- 10 Boxes per Carton
- Recommended to store in cool dry place not more than 35°C

#### Quality Standards

- Manufactured in accordance with ISO 9001 & ISO 13485 Quality Management System and ISO 14001 Environmental Management System
- Exceeds current ASTM 6319 standard for critical defects
- Conforms to ASTM D6978 for use with Chemotherapy Drugs

#### Product Specifications

Thickness Measurements	mm ( $\pm 0.02$ )	mils (Typical)
Finger (15mm from extreme tip)	0.11	4.1
Palm (center of palm)	0.07	2.7
Cuff	0.06	2.4
Length (mm)	$\geq 240$	
Weight by Size (g)	S: 3.2, M: 3.5, L: 3.8, XL: 4.1	

#### Physical Properties

	Before Aging	After Aging
Tensile Strength (Mpa)	$\geq 14$	$\geq 14$
Elongation (%)	500 (min.)	400 (min.)

#### Inspection Levels & AQL

	Inspection Level	AQL
Physical Dimensions	S2	4.0
Physical Properties	S2	4.0
Watertightness	G-1	1.5
Visual Inspection (Major)	G-1	2.5
Visual Inspection (Minor)	G-1	4.0
Powder Residue	$\leq 2\text{mg/glove}$	

#### PULSE MEDICAL FINANCE

The PULSAR™ brand is owned by PULSE Medical Finance, a full-service Supply Chain Platform and FDA Registered Importer. PULSAR™ Nitrile Gloves are manufactured by our network of FDA Registered factories in Malaysia. These are the same factories that produce many of the higher priced glove brands today. Our gloves are tested using the highest international protocol and conform to noted ASTM and ISO standards.



More info at  
[PulsarMedPro.com](http://PulsarMedPro.com)